

**Standard**  
(Sizes >53 cm height 10 only)

**On request**

		width										
Inch		14	15	16	17¼	18	19	20	21	22	23	24
	CM	35	38	40	43	45	48	50	53	55	58	60
14	35	Standard	Standard	Standard	Standard	Standard	On request	On request	On request	On request	On request	On request
15	38	Standard	Standard	Standard	Standard	Standard	On request	On request	On request	On request	On request	On request
16	40	Standard	Standard	Standard	Standard	Standard	Standard	On request	On request	On request	On request	On request
17¼	43	Standard	Standard	Standard	Standard	Standard	Standard	Standard	Standard	On request	On request	On request
18	45	Standard	Standard	Standard	Standard	Standard	Standard	Standard	Standard	Standard	Standard	Standard
19	48	On request	Standard	Standard	Standard	Standard	Standard	Standard	Standard	Standard	Standard	Standard
20	50	On request	On request	Standard	Standard	Standard	Standard	Standard	Standard	Standard	Standard	Standard
21	53	On request	On request	Standard	Standard	Standard	Standard	Standard	Standard	Standard	Standard	Standard
22	55	On request	On request	On request	Standard	Standard	Standard	Standard	Standard	Standard	Standard	Standard
23	58	On request	On request	On request	Standard	Standard	Standard	Standard	Standard	Standard	Standard	Standard
24	60	On request	On request	On request	On request	Standard	Standard	Standard	Standard	Standard	Standard	Standard

## SIZE MEASUREMENT QUICK GUIDE

To get the correct size of wheelchair cushion, Vicair recommends to measure as follows:

The width (X) of the cushion and the wheelchair seat should match. When the cushion rests against armrests or side-guards, the width of a Vicair seat cushion may be up to 2,5 cm (1") wider.\* The depth (Y) of the cushion and seat should match.

\* Exception: The width of the Vicair Active O2 should always match the width of the wheelchair seat.

